

		Inner Diameter		Guidewire compatibility	Microcatheters													
					DraKon™						SeQure®							
					1.7Fr	1.9Fr	2.4Fr	2.7Fr	2.8Fr	3.0Fr	1.7Fr	1.9Fr	2.4Fr	2.7Fr	2.8Fr	3.0Fr		
Angiographic Catheters	4Fr	Terumo Glidecath®	0.041"	1.033mm	0.038"	✓	✓	✓	✓			✓	✓	✓	✓			
		Cordis Tempo®	0.042"	1.067mm	0.038"	✓	✓	✓	✓			✓	✓	✓	✓			
		Merit Impress®	0.040"	1.02mm	0.038"	✓	✓					✓	✓					
		Merit Impress®	0.040"	1.02mm	0.035"													
		BSC Imager™ II Selective	0.039"	0.99mm	0.035"													
	5Fr	Terumo Glidecath®	0.043"	1.10mm	0.038"	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓
		Terumo Glidecath® XP	0.047"	1.20mm	0.038"	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓
		Cordis Tempo®	0.043"	1.09mm	0.038"	✓	✓	✓	✓	✓		✓	✓	✓	✓	✓		
		Merit Impress®	0.046"	1.07mm	0.038"	✓	✓	✓	✓			✓	✓	✓	✓			
		Merit Impress®	0.046"	1.07mm	0.035"													
BSC Imager™ II Selective		0.045"	1.14mm	0.035"														
BSC Imager™ II Selective		0.050"	1.27mm	0.038"	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓

DraKon™ and SeQure® microcatheters are intended for the infusion of contrast media into all peripheral vessels and for drug infusion in intra-arterial therapy, and infusion of embolic materials. They should not be used in cerebral vessels.

CAUTION: Federal law restricts this device to sale by or on the order of a Physician in which the practitioner practices to use or order the use of the device.

DraKon™ and SeQure® IFU.

DraKon™ is a registered trademark of Guerbet Group or its affiliates. SeQure® is a registered trademark of Guerbet Group or its affiliates.

Glidecath® and Glidecath® XP are a registered trademark of Terumo Group or its affiliates.

Tempo® is a registered trademark of Cordis Group or its affiliates.

Impress® is a registered trademark of Merit Group or its affiliates.

Imager™ II Select is a registered trademark of Boston Scientific Group or its affiliates.

GU04210037